## CONFIDENTIAL

BUG M-62

Chief, Operations Division, Medical Staff

Chief, Engineering Staff, OC

	1.	During	the	last	quarter	year,	the	radiologi	cal surve	y and
radi		n detec						were to	ested to	deter-
em ( 69.45	+ ha	fr Aner	at ion	al s	tatua.	fota	1 of	eight ins	eruments:	Were

found to be inoperative. These instruments, with one exception,

were repaired and placed in operating condition.

Status of Radiological Survey and Esdiation Detection Instruments

Cuarterly Report

- 2. One instrument (Thyac type 389, serial No. \_\_\_\_\_ is no longer serviceable because of an unrepairable high voltage power supply. A recommendation is made to cannibalize this unit for its usuable components.
- 3. A request for replacement components and batteries is being prepared for submission early in January 1962. Battery replacements and calibration will be initiated upon receipt of the replacement batteries.
- 4. The total man-hours expended in the last quarter year: 9. The total man-hours expended in 1961: 57.

R&D/Lab/JWM/rkb (9 January 1962)

Distribution:

Original and 1 - Addressee

1 - Lab Subject

1 - OC-E Chrono

1 - Dev/s

1 - Lab Pending

nee	<b>G</b> N		-12-80	008	632
1	glass		. /	-	
DIRO T3UL	FLASS		2010		i
1		ل: بەھىيىد،يىسىسىيىد،	. و سه به المون		

CONFIDENTIAL

25**X**1

25X1

25X1

25X1